

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3349

Name: RITTER 1  
Sterilizer ID: V2178198  
06/07/2023 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
2:00 200.1 F 0.5  
4:00 215.0 F 5.9  
6:00 225.5 F 11.4  
8:00 229.4 F 7.4  
10:00 241.5 F 13.2  
12:00 241.0 F 11.3  
14:00 238.8 F 9.9  
16:00 252.7 F 16.7  
18:00 264.4 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.7 F 29.1  
2:00 271.8 F 28.4  
3:00 271.8 F 28.8  
4:00 272.6 F 29.0  
  
Min 271.3 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 218.3 F 0.0  
10:00 239.7 F 0.0  
15:00 233.6 F 0.0  
20:00 226.3 F 0.0  
25:00 222.3 F 0.0  
30:00 215.9 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 19:58  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:06  
CYCLE COMPLETE

06/07/2023 11:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_