
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004359
11:11 14/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:03
STERILIZING 3:05
136.5 C 223kPa 3:05
136.3 C 225kPa 3:35
136.1 C 218kPa 4:05
136.4 C 225kPa 4:35
136.4 C 222kPa 5:05
136.4 C 222kPa 5:35
136.4 C 223kPa 6:05
136.4 C 225kPa 6:35
136.4 C 225kPa 6:35
VENTING 6:36
AIR DRYING 8:25
CYCLE COMPLETE 15:58

Digital Signature #
AED593D6CFE01D09
