

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3634

Name: RITTER 1  
Sterilizer ID: V2178198  
08/10/2023 11:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.9 F 0.0  
2:00 193.2 F 4.0  
4:00 202.3 F 8.9  
6:00 220.5 F 8.7  
8:00 228.8 F 9.4  
10:00 235.7 F 9.5  
12:00 242.6 F 12.5  
14:00 241.5 F 11.2  
16:00 254.7 F 17.9  
18:00 266.1 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.6 F 29.1  
2:00 271.5 F 28.4  
3:00 272.7 F 29.3  
4:00 272.0 F 28.6  
  
Min 271.5 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 230.5 F 0.0  
10:00 235.9 F 0.0  
15:00 228.5 F 0.0  
20:00 223.0 F 0.0  
25:00 219.8 F 0.0  
30:00 220.4 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 19:37  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:55  
CYCLE COMPLETE

08/10/2023 12:06 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_