
Midmark M11 - v1.0.5
Total Cycles: 2030

Name: RITTER2
Sterilizer ID: V2313228
02/09/2023 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 188.9 F 2.4
4:00 192.1 F 6.8
6:00 212.8 F 12.3
8:00 226.6 F 8.1
10:00 237.5 F 11.7
12:00 241.9 F 12.4
14:00 242.4 F 11.9
16:00 256.2 F 19.1
18:00 268.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 29.4
2:00 271.7 F 29.4
3:00 272.4 F 29.3
4:00 271.7 F 29.3

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 221.6 F 0.0
10:00 228.3 F 0.0
15:00 239.1 F 0.0
20:00 232.8 F 0.0
25:00 226.4 F 0.0
30:00 224.1 F 0.0

Drying Complete

FILLING: 1:33
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

02/09/2023 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____