
Midmark M11 - v1.0.5
Total Cycles: 150

Name: RITTER2
Sterilizer ID: V2313228
06/28/2021 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 198.6 F 4.0
4:00 210.7 F 10.0
6:00 223.3 F 6.8
8:00 238.1 F 12.6
10:00 239.7 F 11.3
12:00 240.7 F 11.0
14:00 254.6 F 18.1
16:00 266.4 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.6 F 29.2
2:00 272.2 F 29.5
3:00 272.2 F 29.0
4:00 272.4 F 29.7

Min 271.6 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 216.8 F 0.0
10:00 231.1 F 0.0
15:00 232.7 F 0.0
20:00 226.8 F 0.0
25:00 219.3 F 0.0
30:00 215.2 F 0.0
Drying Complete

FILLING: 2:50
HEATING: 17:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:32
CYCLE COMPLETE

06/28/2021 02:14 PM
ACCEPTED: X
REJECTED:
Operator: U01