
Midmark M11 - v1.0.5
Total Cycles: 4099

Name: RITTER2
Sterilizer ID: V2313228
11/07/2024 05:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 127.7 F 0.0
4:00 155.4 F 0.7
6:00 207.4 F 4.4
8:00 202.5 F 8.6
10:00 215.1 F 12.7
12:00 229.0 F 7.3
14:00 240.1 F 12.1
16:00 234.8 F 9.4
18:00 238.3 F 10.0
20:00 247.1 F 14.0
22:00 258.7 F 20.4
24:00 268.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.4
1:00 272.7 F 29.4
2:00 272.6 F 29.5
3:00 272.9 F 29.6
4:00 272.9 F 29.7

Min 271.5 F 29.3
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.4 F 0.0
5:00 236.1 F 0.0
10:00 184.9 F 0.0
15:00 179.0 F 0.0
20:00 176.5 F 0.0
25:00 174.7 F 0.0
30:00 176.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:35
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:24
CYCLE COMPLETE

11/07/2024 06:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____