
Midmark M11 - v1.0.0
Total Cycles: 4195

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2024 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.1 F 0.0
2:00 128.1 F 1.4
4:00 150.0 F 3.7
6:00 177.8 F 7.3
8:00 198.5 F 11.2
10:00 214.4 F 3.6
12:00 221.6 F 7.3
14:00 230.5 F 11.3
16:00 230.3 F 6.9
18:00 241.9 F 12.3
20:00 239.5 F 10.4
22:00 251.8 F 16.3
24:00 262.0 F 22.3
26:00 271.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.8 F 29.1
2:00 272.5 F 29.1
3:00 272.7 F 29.2
4:00 272.4 F 29.3

Min 271.0 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 232.2 F 0.0
10:00 224.1 F 0.0
15:00 237.4 F 0.0
20:00 231.4 F 0.0
25:00 217.7 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 26:24
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:05:15
CYCLE COMPLETE

01/11/2024 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____