
Midmark M11 - v1.0.0
Total Cycles: 1416

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2021 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.9 F 0.0
2:00 143.8 F 1.3
4:00 163.0 F 4.1
6:00 190.6 F 8.4
8:00 215.7 F 12.7
10:00 226.5 F 7.5
12:00 238.5 F 12.4
14:00 236.2 F 9.1
16:00 238.7 F 9.7
18:00 247.0 F 13.6
20:00 259.2 F 20.2
22:00 270.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.8 F 29.0
2:00 272.0 F 28.4
3:00 271.6 F 28.6
4:00 272.7 F 29.0

Min 271.0 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 223.6 F 0.0
10:00 218.9 F 0.0
15:00 231.1 F 0.0
20:00 223.6 F 0.0
25:00 228.3 F 0.0
30:00 195.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:55
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

12/01/2021 02:40 PM
ACCEPTED: X
REJECTED:
Operator: U02