
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001889
9:49 25/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 2:59
136.5 C 220kPa 2:59
136.4 C 226kPa 3:29
136.4 C 225kPa 3:59
136.3 C 225kPa 4:29
136.3 C 224kPa 4:59
136.3 C 225kPa 5:29
136.1 C 223kPa 5:59
136.2 C 224kPa 6:29
136.2 C 224kPa 6:29
VENTING 6:29
AIR DRYING 7:17
CYCLE COMPLETE 12:32

Digital Signature #
2C676F6098C8EF37
