

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001685  
10:23 24/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:09  
STERILIZING 4:07  
136.0 C 220kPa 4:07  
136.4 C 224kPa 4:37  
136.4 C 224kPa 5:07  
136.2 C 222kPa 5:37  
136.3 C 223kPa 6:07  
136.4 C 224kPa 6:37  
136.3 C 221kPa 7:07  
136.4 C 223kPa 7:37  
136.4 C 223kPa 7:37  
VENTING 7:37  
AIR DRYING 8:25  
CYCLE COMPLETE 9:16  
-----

Digital Signature #  
E8F12EDE1D64E154  
-----