
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001465
12:34 06/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:59
PRESSURIZING 3:59
STERILIZING 5:01
136.4 C 223kPa 5:01
136.3 C 222kPa 5:31
136.2 C 224kPa 6:01
136.3 C 224kPa 6:31
136.4 C 225kPa 7:01
136.4 C 224kPa 7:31
136.3 C 221kPa 8:01
136.1 C 221kPa 8:31
136.1 C 221kPa 8:31
VENTING 8:31
AIR DRYING 9:29
CYCLE COMPLETE 9:41

Digital Signature #
D646E97B8D0A7C56
