
Midmark M11 - v1.0.5
Total Cycles: 854

Name: RITTER2
Sterilizer ID: V2313228
03/01/2022 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.0 F 0.0
2:00 155.7 F 3.4
4:00 182.5 F 7.8
6:00 206.2 F 13.4
8:00 221.6 F 7.5
10:00 237.7 F 13.2
12:00 239.4 F 11.1
14:00 239.7 F 10.6
16:00 254.9 F 18.3
18:00 267.4 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.1
1:00 272.1 F 29.0
2:00 272.6 F 29.5
3:00 272.6 F 29.8
4:00 272.2 F 29.0

Min 271.8 F 28.9
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 213.7 F 0.0
10:00 207.4 F 0.0
15:00 211.6 F 0.0
20:00 212.7 F 0.0
25:00 213.9 F 0.0
30:00 211.3 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 19:15
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:57:42
CYCLE COMPLETE

03/01/2022 05:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____