
Midmark M11 - v1.0.5
Total Cycles: 177

Name: RITTER2
Sterilizer ID: V2313228
07/07/2021 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.4 F 0.0
2:00 195.5 F 4.1
4:00 208.4 F 9.9
6:00 222.8 F 7.4
8:00 235.3 F 12.1
10:00 236.9 F 10.5
12:00 237.1 F 9.5
14:00 251.0 F 16.3
16:00 262.5 F 23.1
18:00 272.9 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.7 F 29.7
2:00 271.7 F 29.2
3:00 272.7 F 29.8
4:00 271.8 F 29.1

Min 271.5 F 29.1
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 227.8 F 0.0
10:00 230.4 F 0.0
15:00 228.6 F 0.0
20:00 197.1 F 0.0
25:00 187.0 F 0.0
Dry Cycle Aborted

FILLING: 2:60
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:11
DRYING: 28:11
TOTAL CYCLE: 00:56:40

07/07/2021 04:13 PM
ACCEPTED: X
REJECTED:
Operator: U01