
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001586
11:20 13/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:25
STERILIZING 3:31
136.5 C 224kPa 3:31
136.4 C 226kPa 4:01
136.4 C 224kPa 4:31
136.3 C 225kPa 5:01
136.1 C 223kPa 5:31
136.4 C 226kPa 6:01
136.4 C 225kPa 6:31
136.3 C 224kPa 7:01
136.3 C 224kPa 7:01
VENTING 7:01
AIR DRYING 7:53
CYCLE COMPLETE 8:00

Digital Signature #
BA52307D559C95A2
