
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003080
8:54 03/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:17
STERILIZING 4:22
136.4 C 222kPa 4:22
136.4 C 222kPa 4:52
136.0 C 218kPa 5:22
136.4 C 224kPa 5:52
136.0 C 218kPa 6:22
136.4 C 223kPa 6:52
136.4 C 222kPa 7:22
136.4 C 221kPa 7:52
136.4 C 221kPa 7:52
VENTING 7:53
AIR DRYING 8:46
CYCLE COMPLETE 8:50

Digital Signature #
66860633C0B87EEE
