

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2036

Name: RITTER 1  
Sterilizer ID: V2178198  
06/20/2022 12:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.6 F 0.0  
2:00 135.1 F 0.3  
4:00 163.4 F 2.0  
6:00 185.6 F 4.9  
8:00 213.4 F 9.0  
10:00 227.9 F 12.5  
12:00 227.4 F 7.7  
14:00 236.7 F 12.4  
16:00 235.0 F 8.7  
18:00 241.6 F 11.2  
20:00 244.1 F 12.2  
22:00 256.1 F 18.4  
24:00 266.5 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.1 F 29.1  
2:00 272.1 F 28.5  
3:00 271.1 F 28.2  
4:00 272.0 F 29.1

Min 271.1 F 28.1  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 192.0 F 0.0  
10:00 202.3 F 0.0  
15:00 209.1 F 0.0  
20:00 210.8 F 0.0  
25:00 227.5 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 25:43  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:03:43  
CYCLE COMPLETE

06/20/2022 01:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_