
Midmark M11 - v1.0.5
Total Cycles: 4448

Name: RITTER2
Sterilizer ID: V2313228
03/24/2025 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 146.4 F 2.0
4:00 159.4 F 5.4
6:00 189.0 F 9.6
8:00 210.6 F 8.1
10:00 224.3 F 8.5
12:00 236.7 F 13.3
14:00 235.5 F 10.5
16:00 236.6 F 8.5
18:00 248.2 F 15.3
20:00 259.8 F 21.8
22:00 270.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.7
1:00 272.6 F 30.5
2:00 272.6 F 30.6
3:00 272.6 F 30.8
4:00 271.8 F 30.4

Min 271.0 F 29.7
Max 273.1 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 201.7 F 0.0
10:00 228.2 F 0.0
15:00 224.8 F 0.0
20:00 223.3 F 0.0
25:00 222.1 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:52
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:01:47
CYCLE COMPLETE

03/24/2025 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____