
Midmark M11 - v1.0.0
Total Cycles: 3501

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2023 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	166.9 F	0.0
2:00	164.7 F	0.4
4:00	174.9 F	3.6
6:00	199.4 F	8.8
8:00	219.3 F	7.1
10:00	234.0 F	10.5
12:00	233.9 F	8.1
14:00	239.4 F	9.7
16:00	248.5 F	14.3
18:00	261.9 F	21.7
20:00	272.2 F	28.2

C534: STERILIZE MODE
STEAM TEMP LOW
CHAMBER TEMP 270F
CHAMBER PRESS 21.9PSI
ELAPSED TIME 00:01
** ITEMS NOT STERILE **

07/07/2023 10:55 AM