
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001951
10:20 14/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:03
STERILIZING 4:10
136.0 C 221kPa 4:10
136.4 C 225kPa 4:40
136.3 C 223kPa 5:10
136.3 C 224kPa 5:40
136.4 C 225kPa 6:10
136.3 C 224kPa 6:40
136.2 C 224kPa 7:10
136.4 C 225kPa 7:40
136.4 C 225kPa 7:40
VENTING 7:40
AIR DRYING 8:30
CYCLE COMPLETE 22:40

Digital Signature #
BDA4AFB665ABCD9B
