

-----  
Midmark M11 - v1.0.5  
Total Cycles: 856

Name: RITTER2  
Sterilizer ID: V2313228  
03/02/2022 09:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.2 F 0.0  
2:00 122.0 F 1.2  
4:00 146.3 F 3.5  
6:00 169.0 F 7.3  
8:00 190.2 F 11.9  
10:00 210.3 F 5.2  
12:00 223.1 F 10.0  
14:00 230.2 F 8.4  
16:00 239.0 F 12.2  
18:00 238.8 F 10.8  
20:00 252.3 F 17.4  
22:00 263.7 F 24.1  
24:00 271.9 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.7 F 30.2  
2:00 271.6 F 29.4  
3:00 271.6 F 30.0  
4:00 272.7 F 30.1

Min 271.1 F 29.3  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
5:00 200.2 F 0.0  
10:00 210.5 F 0.0  
15:00 215.3 F 0.0  
20:00 217.1 F 0.0  
25:00 216.9 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 24:09  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:02:13  
CYCLE COMPLETE

03/02/2022 10:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_