
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 002543
10:02 27/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:40
STERILIZING 3:39
136.4 C 224kPa 3:39
136.4 C 225kPa 4:09
136.1 C 222kPa 4:39
136.4 C 224kPa 5:09
136.3 C 223kPa 5:39
136.1 C 222kPa 6:09
136.1 C 220kPa 6:39
136.4 C 221kPa 7:09
136.4 C 221kPa 7:09
VENTING 7:09
AIR DRYING 8:01
CYCLE COMPLETE 8:02

Digital Signature #
D423096E11181D2F
