

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5458

Name: RITTER 1  
Sterilizer ID: V2178198  
03/27/2025 04:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.9 F 0.0  
2:00 179.4 F 3.7  
4:00 189.1 F 8.3  
6:00 214.1 F 10.5  
8:00 221.6 F 8.1  
10:00 235.8 F 12.3  
12:00 238.2 F 11.2  
14:00 237.3 F 9.8  
16:00 252.0 F 16.8  
18:00 264.3 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 272.3 F 29.6  
2:00 271.8 F 28.9  
3:00 271.9 F 29.3  
4:00 272.1 F 29.2  
  
Min 271.3 F 28.7  
Max 272.4 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 200.3 F 0.0  
10:00 208.6 F 0.0  
15:00 210.9 F 0.0  
20:00 212.7 F 0.0  
25:00 218.6 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 00:58:44  
CYCLE COMPLETE

03/27/2025 05:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_