
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001477
13:44 10/06/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 5:52
STERILIZING 7:32
136.4 C 223kPa 7:32
136.4 C 223kPa 8:02
136.3 C 222kPa 8:32
136.5 C 223kPa 9:03
136.2 C 221kPa 9:33
136.4 C 223kPa 10:03
136.4 C 221kPa 10:33
136.4 C 219kPa 11:02
VENTING 11:03
AIR DRYING 12:11
CYCLE COMPLETE 28:45

Digital Signature #
E18CFCA1FAEFC16C
