
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000238
10:40 20/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:14
STERILIZING 4:36
136.4 C 220kPa 4:36
136.3 C 223kPa 5:06
136.3 C 224kPa 5:36
136.2 C 219kPa 6:06
136.2 C 223kPa 6:36
136.2 C 222kPa 7:06
136.4 C 223kPa 7:36
136.2 C 223kPa 8:06
136.2 C 223kPa 8:06
VENTING 8:06
AIR DRYING 9:02
CYCLE COMPLETE 9:07

Digital Signature #
3994CE6D9ACFE173
