
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004968
15:30 19/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:13
STERILIZING 4:22
136.4 C 223kPa 4:22
136.2 C 221kPa 4:52
136.3 C 217kPa 5:22
136.2 C 221kPa 5:52
136.3 C 223kPa 6:22
136.3 C 222kPa 6:52
136.4 C 221kPa 7:22
136.4 C 223kPa 7:52
136.4 C 223kPa 7:52
VENTING 7:52
AIR DRYING 8:38
CYCLE COMPLETE 8:59

Digital Signature #
15E50F5089D935C4
