

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002607  
7:17 03/11/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:04  
STERILIZING 3:06  
136.5 C 223kPa 3:06  
136.3 C 223kPa 3:36  
136.4 C 223kPa 4:06  
136.0 C 221kPa 4:36  
136.1 C 221kPa 5:06  
136.1 C 221kPa 5:36  
136.0 C 219kPa 6:06  
136.3 C 222kPa 6:36  
136.3 C 222kPa 6:36  
VENTING 6:36  
AIR DRYING 7:24  
CYCLE COMPLETE 7:31

-----  
Digital Signature #  
39D335543D72EBF4  
-----