
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005941
8:46 11/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:28
STERILIZING 3:22
136.0 C 220kPa 3:22
136.2 C 224kPa 3:52
135.9 C 221kPa 4:22
136.4 C 223kPa 4:52
136.3 C 224kPa 5:22
136.4 C 225kPa 5:52
136.4 C 223kPa 6:22
136.4 C 223kPa 6:52
136.4 C 223kPa 6:52
VENTING 6:52
AIR DRYING 7:42
CYCLE COMPLETE 7:53

Digital Signature #
63F5FFEB7ECCADDD
