
Midmark M11 - v1.0.0
Total Cycles: 3150

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2023 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.4 F 0.0
2:00 175.1 F 0.0
4:00 189.6 F 1.5
6:00 207.4 F 6.3
8:00 226.4 F 11.3
10:00 230.1 F 7.1
12:00 240.8 F 12.8
14:00 239.2 F 10.5
16:00 237.4 F 8.8
18:00 251.3 F 15.8
20:00 263.1 F 22.7
22:00 272.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 271.7 F 28.3
3:00 272.3 F 28.9
4:00 272.5 F 28.8

Min 271.4 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 212.5 F 0.0
10:00 221.1 F 0.0
15:00 227.9 F 0.0
20:00 227.2 F 0.0
25:00 226.1 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

04/24/2023 10:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____