

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1590

Name: RITTER2  
Sterilizer ID: V2313228  
10/13/2022 06:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.9 F 0.0  
2:00 81.8 F 0.0  
4:00 97.3 F 0.0  
6:00 115.1 F 0.0  
8:00 162.6 F 0.1  
10:00 187.1 F 0.3  
12:00 206.7 F 3.5  
14:00 218.6 F 6.8  
16:00 227.3 F 10.3  
18:00 233.9 F 9.8  
20:00 235.1 F 9.9  
22:00 240.8 F 9.8  
24:00 243.3 F 13.0  
26:00 241.8 F 11.8  
28:00 253.8 F 18.0  
30:00 264.4 F 24.4  
32:00 271.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 272.6 F 29.6  
2:00 272.3 F 29.8  
3:00 272.2 F 29.3  
4:00 272.9 F 30.0  
  
Min 271.7 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 227.3 F 0.0  
10:00 222.1 F 0.0  
15:00 221.2 F 0.0  
20:00 212.9 F 0.0  
25:00 212.0 F 0.0  
30:00 208.4 F 0.0  
  
Drying Complete

FILLING: 1:03  
HEATING: 32:02  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:10:13  
CYCLE COMPLETE

10/13/2022 07:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_