

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2269

Name: RITTER2  
Sterilizer ID: V2313228  
04/19/2023 01:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
2:00 179.3 F 3.5  
4:00 198.4 F 8.4  
6:00 218.5 F 9.8  
8:00 229.6 F 9.9  
10:00 233.3 F 8.3  
12:00 241.5 F 11.4  
14:00 246.6 F 14.0  
16:00 259.6 F 21.2  
18:00 271.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.4 F 29.3  
2:00 272.1 F 29.7  
3:00 271.9 F 29.0  
4:00 272.6 F 29.7  
  
Min 271.4 F 28.8  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.9 F 0.0  
5:00 215.9 F 0.0  
10:00 211.5 F 0.0  
15:00 233.3 F 0.0  
20:00 225.6 F 0.0  
25:00 225.5 F 0.0  
30:00 220.5 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 18:36  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

04/19/2023 02:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_