
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001334
9:58 09/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:27
PRESSURIZING 3:31
STERILIZING 4:37
136.0 C 220kPa 4:37
136.2 C 223kPa 5:07
136.2 C 223kPa 5:37
136.4 C 224kPa 6:07
136.1 C 221kPa 6:37
136.5 C 225kPa 7:07
136.4 C 223kPa 7:37
136.3 C 223kPa 8:07
136.3 C 223kPa 8:07
VENTING 8:08
AIR DRYING 9:02
CYCLE COMPLETE 13:31

Digital Signature #
D2AFE6CF8A92FFA2
