
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001051
14:14 29/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:58
STERILIZING 4:07
136.0 C 222kPa 4:07
136.1 C 222kPa 4:37
136.4 C 223kPa 5:07
136.0 C 221kPa 5:37
136.1 C 223kPa 6:07
136.4 C 225kPa 6:37
136.2 C 224kPa 7:07
136.4 C 225kPa 7:37
136.4 C 225kPa 7:37
VENTING 7:37
AIR DRYING 8:27
CYCLE COMPLETE 20:09

Digital Signature #
E8473D0D20D3E724
