

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1787

Name: RITTER 1  
Sterilizer ID: V2178198  
04/04/2022 08:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.8 F 0.0  
2:00 166.9 F 1.4  
4:00 175.8 F 4.9  
6:00 190.5 F 9.1  
8:00 211.6 F 9.4  
10:00 219.9 F 7.4  
12:00 234.6 F 12.2  
14:00 234.5 F 8.8  
16:00 240.8 F 10.7  
18:00 244.9 F 12.8  
20:00 256.8 F 19.1  
22:00 267.8 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.7 F 28.9  
2:00 271.2 F 28.4  
3:00 272.3 F 28.9  
4:00 271.2 F 28.2

Min 271.1 F 28.2  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 229.9 F 0.0  
10:00 232.2 F 0.0  
15:00 228.7 F 0.0  
20:00 225.2 F 0.0  
25:00 226.8 F 0.0  
30:00 221.3 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 23:22  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:01:24  
CYCLE COMPLETE

04/04/2022 09:35 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02