

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001073  
10:14 12/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:09  
PRESSURIZING 4:10  
STERILIZING 5:30  
136.0 C 221kPa 5:30  
136.2 C 223kPa 6:01  
136.4 C 224kPa 6:31  
136.3 C 224kPa 7:01  
136.4 C 224kPa 7:31  
136.4 C 224kPa 8:01  
136.4 C 224kPa 8:31  
136.3 C 224kPa 9:01  
136.3 C 224kPa 9:01  
VENTING 9:01  
AIR DRYING 9:59  
CYCLE COMPLETE 24:59  
-----

Digital Signature #  
5CE114AA330D6FFC  
-----