

-----  
Midmark M11 - v1.0.5  
Total Cycles: 345

Name: RITTER2  
Sterilizer ID: V2313228  
09/09/2021 09:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
2:00 184.3 F 3.6  
4:00 206.3 F 9.3  
6:00 223.2 F 7.2  
8:00 236.1 F 12.2  
10:00 238.1 F 10.8  
12:00 238.1 F 10.2  
14:00 253.1 F 17.5  
16:00 265.1 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.2 F 29.3  
2:00 272.7 F 29.9  
3:00 271.8 F 29.1  
4:00 272.4 F 29.6  
  
Min 271.6 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 241.8 F 0.0  
10:00 217.0 F 0.0  
15:00 223.0 F 0.0  
20:00 213.0 F 0.0  
25:00 196.1 F 0.0  
30:00 203.3 F 0.0  
Drying Complete

FILLING: 3:31  
HEATING: 17:44  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

09/09/2021 10:49 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02