
Midmark M11 - v1.0.0
Total Cycles: 822

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2021 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.4 F 0.0
2:00 138.8 F 1.4
4:00 165.1 F 4.1
6:00 188.9 F 8.4
8:00 209.7 F 13.0
10:00 223.2 F 7.1
12:00 237.3 F 12.3
14:00 235.7 F 9.1
16:00 239.1 F 9.6
18:00 247.3 F 13.7
20:00 259.5 F 20.4
22:00 270.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.6
2:00 271.7 F 28.5
3:00 272.2 F 28.4
4:00 271.9 F 28.6

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 220.9 F 0.0
10:00 240.1 F 0.0
15:00 227.1 F 0.0
20:00 219.5 F 0.0
25:00 222.0 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

05/19/2021 12:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____