
Midmark M11 - v1.0.0
Total Cycles: 1254

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2021 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.8 F 0.0
2:00 158.1 F 2.1
4:00 170.9 F 5.4
6:00 195.0 F 9.4
8:00 215.9 F 8.6
10:00 225.7 F 8.8
12:00 237.6 F 12.1
14:00 238.4 F 10.9
16:00 235.8 F 8.7
18:00 249.4 F 15.1
20:00 260.5 F 21.3
22:00 270.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.6 F 29.1
2:00 271.6 F 28.3
3:00 272.7 F 29.2
4:00 272.0 F 28.6

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 183.9 F 0.0
10:00 187.5 F 0.0
15:00 194.1 F 0.0
20:00 196.7 F 0.0
25:00 228.9 F 0.0
30:00 196.0 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:49
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

10/12/2021 11:06 AM
ACCEPTED: X
REJECTED:
Operator: U01