
Midmark M11 - v1.0.0
Total Cycles: 4346

Name: RITTER 1
Sterilizer ID: V2178198
02/29/2024 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.9 F 0.0
2:00 158.1 F 2.8
4:00 170.8 F 6.1
6:00 196.1 F 10.9
8:00 216.0 F 4.1
10:00 228.9 F 9.0
12:00 236.9 F 10.9
14:00 240.0 F 10.6
16:00 235.3 F 8.3
18:00 250.1 F 15.1
20:00 262.1 F 21.9
22:00 272.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.7 F 28.8
2:00 271.5 F 28.1
3:00 272.1 F 28.8
4:00 272.3 F 28.6

Min 271.0 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 245.1 F 0.0
10:00 221.0 F 0.0
15:00 231.7 F 0.0
20:00 232.7 F 0.0
25:00 232.6 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

02/29/2024 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____