
Midmark M11 - v1.0.0
Total Cycles: 957

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2021 08:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.5 F 0.0
2:00 196.2 F 2.0
4:00 206.5 F 5.6
6:00 217.3 F 10.6
8:00 223.8 F 5.1
10:00 232.2 F 9.6
12:00 239.8 F 10.0
14:00 241.7 F 11.9
16:00 239.6 F 10.2
18:00 252.8 F 16.8
20:00 264.1 F 23.4
22:00 271.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.8
2:00 271.8 F 28.8
3:00 272.0 F 28.6
4:00 272.0 F 28.9

Min 271.4 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 225.2 F 0.0
10:00 232.6 F 0.0
15:00 220.7 F 0.0
20:00 214.8 F 0.0
25:00 206.6 F 0.0
30:00 199.0 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

07/05/2021 09:06 AM
ACCEPTED: X
REJECTED:
Operator: U01