
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000819
14:11 10/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:19
STERILIZING 4:25
136.0 C 228kPa 4:25
136.3 C 232kPa 4:55
136.3 C 232kPa 5:25
136.3 C 232kPa 5:55
136.4 C 233kPa 6:25
136.4 C 233kPa 6:55
136.4 C 233kPa 7:25
136.4 C 233kPa 7:55
136.4 C 233kPa 7:55
VENTING 7:55
AIR DRYING 8:49
CYCLE COMPLETE 23:50

Digital Signature #
A59A7BB0C67F8395
