

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001069  
10:00 10/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:25  
STERILIZING 5:13  
136.4 C 221kPa 5:13  
136.3 C 223kPa 5:43  
136.2 C 223kPa 6:13  
136.3 C 223kPa 6:43  
136.4 C 225kPa 7:13  
136.2 C 223kPa 7:43  
136.4 C 224kPa 8:13  
136.3 C 224kPa 8:43  
136.3 C 224kPa 8:43  
VENTING 8:43  
AIR DRYING 9:42  
CYCLE COMPLETE 24:42

-----  
Digital Signature #  
C22492731FF1E4D8  
-----