
Midmark M11 - v1.0.5
Total Cycles: 564

Name: RITTER2
Sterilizer ID: V2313228
11/18/2021 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.6 F 0.0
2:00 178.3 F 2.6
4:00 183.9 F 7.3
6:00 205.9 F 12.7
8:00 219.3 F 7.6
10:00 232.7 F 12.9
12:00 236.5 F 10.4
14:00 235.8 F 8.9
16:00 249.8 F 15.9
18:00 261.9 F 22.8
20:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.7 F 29.6
2:00 271.6 F 29.7
3:00 272.2 F 29.4
4:00 271.8 F 29.8

Min 271.2 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 213.4 F 0.0
10:00 192.4 F 0.0
15:00 191.6 F 0.0
20:00 188.3 F 0.0
25:00 184.2 F 0.0
30:00 182.7 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

11/18/2021 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____