
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000638
8:25 09/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:37
STERILIZING 3:45
136.0 C 220kPa 3:45
136.4 C 233kPa 4:15
136.4 C 229kPa 4:45
136.2 C 230kPa 5:15
136.1 C 230kPa 5:45
136.4 C 231kPa 6:15
136.4 C 231kPa 6:45
136.2 C 230kPa 7:15
136.2 C 230kPa 7:15
VENTING 7:15
AIR DRYING 8:06
CYCLE COMPLETE 23:06

Digital Signature #
B0736F49CD35E5D0
