
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 000845
9:34 20/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:44
STERILIZING 4:53
136.0 C 220kPa 4:53
136.3 C 223kPa 5:23
136.3 C 221kPa 5:53
136.1 C 221kPa 6:23
136.2 C 223kPa 6:53
135.9 C 221kPa 7:23
136.4 C 223kPa 7:53
136.2 C 221kPa 8:23
136.2 C 221kPa 8:23
VENTING 8:23
AIR DRYING 9:19
CYCLE COMPLETE 9:25

Digital Signature #
D9365E6F2FEAD5A1
