
Midmark M11 - v1.0.5
Total Cycles: 975

Name: RITTER2
Sterilizer ID: V2313228
04/11/2022 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 191.8 F 2.0
4:00 203.9 F 7.7
6:00 222.6 F 13.4
8:00 232.3 F 10.3
10:00 234.5 F 8.6
12:00 239.4 F 10.0
14:00 249.7 F 15.6
16:00 262.6 F 23.1
18:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.4 F 29.6
2:00 271.6 F 29.4
3:00 272.5 F 29.5
4:00 271.3 F 29.2

Min 271.3 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 211.9 F 0.0
10:00 219.9 F 0.0
15:00 226.0 F 0.0
20:00 235.0 F 0.0
25:00 229.4 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:07
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:14
CYCLE COMPLETE

04/11/2022 01:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____