
Midmark M11 - v1.0.0
Total Cycles: 1785

Name: RITTER 1
Sterilizer ID: V2178198
04/01/2022 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.8 F 0.0
2:00 191.3 F 4.3
4:00 217.8 F 9.9
6:00 223.5 F 6.0
8:00 239.2 F 11.6
10:00 235.3 F 9.3
12:00 237.4 F 9.0
14:00 248.9 F 15.2
16:00 261.9 F 22.4
18:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.1 F 28.9
2:00 272.4 F 29.0
3:00 272.5 F 29.3
4:00 272.5 F 29.5

Min 271.5 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 232.8 F 0.0
10:00 229.1 F 0.0
15:00 226.4 F 0.0
20:00 220.4 F 0.0
25:00 223.2 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 2:55
HEATING: 18:07
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

04/04/2022 07:13 AM
ACCEPTED: X
REJECTED:
Operator: U02