
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000818
13:08 10/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 3:04
STERILIZING 4:04
136.0 C 227kPa 4:04
136.5 C 233kPa 4:34
136.2 C 230kPa 5:04
136.1 C 230kPa 5:34
136.4 C 232kPa 6:04
136.4 C 232kPa 6:34
136.4 C 232kPa 7:04
136.2 C 228kPa 7:34
136.2 C 228kPa 7:34
VENTING 7:34
AIR DRYING 8:26
CYCLE COMPLETE 23:26

Digital Signature #
6E493FF4612B8459
