

-----  
Midmark M11 - v1.0.0  
Total Cycles: 910

Name: RITTER 1  
Sterilizer ID: V2178198  
06/17/2021 11:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 116.3 F 0.0  
2:00 116.1 F 0.0  
4:00 127.2 F 0.1  
6:00 160.9 F 0.1  
8:00 191.0 F 2.5  
10:00 205.9 F 5.2  
12:00 212.2 F 8.1  
14:00 221.7 F 11.3  
16:00 229.8 F 6.0  
18:00 232.3 F 9.8  
20:00 239.7 F 10.9  
22:00 242.3 F 11.2  
24:00 237.1 F 8.9  
26:00 249.3 F 14.6  
28:00 260.1 F 20.5  
30:00 270.0 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.9  
1:00 272.5 F 28.9  
2:00 272.4 F 28.9  
3:00 271.9 F 28.8  
4:00 271.8 F 28.4

Min 270.8 F 27.8  
Max 273.0 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.0 F 0.0  
5:00 241.8 F 0.0  
10:00 239.4 F 0.0  
15:00 227.3 F 0.0  
20:00 220.0 F 0.0  
25:00 219.5 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 30:53  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:09:30  
CYCLE COMPLETE

06/17/2021 12:21 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_