
Midmark M11 - v1.0.5
Total Cycles: 1885

Name: RITTER2
Sterilizer ID: V2313228
12/28/2022 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.1 F 0.0
2:00 117.4 F 0.0
4:00 129.0 F 0.0
6:00 166.6 F 0.3
8:00 192.6 F 4.1
10:00 209.4 F 8.2
12:00 223.3 F 12.8
14:00 229.9 F 8.4
16:00 240.9 F 12.3
18:00 241.6 F 11.8
20:00 240.5 F 10.8
22:00 254.0 F 17.6
24:00 265.7 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.5
2:00 271.7 F 28.8
3:00 271.5 F 29.1
4:00 272.5 F 29.4

Min 271.1 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 211.7 F 0.0
10:00 224.6 F 0.0
15:00 226.8 F 0.0
20:00 225.7 F 0.0
25:00 225.9 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:44
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

12/28/2022 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____