
Midmark M11 - v1.0.5
Total Cycles: 2583

Name: RITTER2
Sterilizer ID: V2313228
06/28/2023 02:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.5 F 0.0
2:00 166.2 F 2.9
4:00 184.3 F 7.3
6:00 212.7 F 13.1
8:00 222.4 F 8.3
10:00 235.0 F 11.4
12:00 239.1 F 11.6
14:00 238.7 F 10.3
16:00 252.8 F 17.3
18:00 264.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.7 F 29.7
2:00 271.6 F 28.9
3:00 272.6 F 29.9
4:00 272.3 F 29.4

Min 271.5 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 215.0 F 0.0
10:00 224.9 F 0.0
15:00 226.1 F 0.0
20:00 228.3 F 0.0
25:00 227.9 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

06/28/2023 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____